



Ref. Certif. No.

SG-MD-00422M1

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)
CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE CERTIFICAT D'ESSAI OC

Product
Produit

Switching power supply unit
(Switching Power Adaptor)

Name and address of the applicant
Nom et adresse du demandeur

XP Power Limited
401 Commonwealth Drive,
Haw Par Technocentre, Lobby B, #02-02 149598, SINGAPORE

Name and address of the manufacturer
Nom et adresse du fabricant

XP Power Limited, 401 Commonwealth Drive,, Haw Par Technocentre,
Lobby B, #02-02 149598, SINGAPORE

Name and address of the factory
Nom et adresse de l'usine

Xianglong Road, Huangzhou, New Town District, Shilong Town, 523326
Dongguan, Guangdong Province, PEOPLE'S REPUBLIC OF CHINA

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

Rated Input : 100-240VAC, 50-60Hz, 1.1A
Rated Output : See test report for details
Protection Class : I (for models AFM30USZZC6,
AFM30USZZ);
II (for models AFM30USZZC2)
Degree of Protection : IP20

Trade mark (if any)
Marque de fabrique (si elle existe)

XP

Model/type Ref.
Ref. de type

AFM30USZZC2, AFM30USZZC6, AFM30USZZ
(See test report for details of model description)

Additional information (if necessary)
Information complémentaire (si nécessaire)

Certificate SG-MD-00422 issued 2014-03-31 is replaced by this version
due to technical changes.

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60601-1(ed.3);am1

as shown in the Test Report Ref. No.
which forms part of this certificate
comme indiqué dans le Rapport d'essais numéro
de référence qui constitue une partie de ce
certificat

TÜV SÜD PSB Pte Ltd
211-400541-100

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

Date, 2015-04-08
CBS 15 03 65969 127

(Yager Bi)




TÜV SÜD PSB Pte Ltd · 1 Science Park Drive · Singapore 118221

PSB Singapore

Additional information:

*The risk management requirements of the standard
were not addressed.*



Date, 2015-04-08
CBS 15 03 65969 127

Signature : (Yager Bi)

TÜV SÜD PSB Pte Ltd • 1 Science Park Drive • Singapore 118221



PSB Singapore